



2931 Soldier Springs Road
Laramie, WY 82070
307.742.5452
www.rt-corp.com

Asbestos TEM Laboratories

ATTN: Marisa Diaz
1409 Fifth Street, Ste C
Berkeley, CA 94710
US

EPA Lab CA00409

PHONE 510-528-0108

FAX 510-528-0109

EMAIL mark@asbestostemplabs.com



This report may contain data that are not covered by the NVLAP accreditation.

PEI-016-1

Trace Metals 1

Program: WSCHEM

Evaluation

Analyte	Result Units	Method	Method ID	Evaluation	Z
1075 Lead	28.4 µg/L	EPA 200.9	10015006	Acceptable	-0.795

PEI-016-1

Study Summary

Analyte	Units	EPA Mean	Assigned Value	Study Mean	Study Std. Dev.	Acceptance Limits	Warning Limits
1075 Lead	µg/L	31.0	31.0	30.7	3.27	21.7 - 40.3	

PEI-016-1

PEI-081

Asbestos

Program: WSCHEM

Evaluation

Analyte	Result Units	Method	Method ID	Evaluation	Z
1520 Asbestos	6.77 MF/L	EPA 100.2	10004407	Acceptable	-0.0535

PEI-081

Study Summary

Analyte	Units	EPA Mean	Assigned Value	Study Mean	Study Std. Dev.	Acceptance Limits	Warning Limits
1520 Asbestos	MF/L	7.00	7.00	7.00	1.26	0.000 - 15.6	

PEI-081

Authorized for release by

Certifying Officer - QA/QC Manager

Date 12/27/2004

Questions / Comments?

Christopher Rucinski

phone: (307) 742-5452

email: reports@rt-corp.com

This report shall not be reproduced, except in full, without the written approval of the laboratory. A laboratory may not claim endorsement by NVLAP, NIST, or any other federal agency. RTC is accredited by NVLAP to perform PT programs for the scope of accreditation under NVLAP Lab Code 200393-0.

Blank
Page

End of Report

Study **WS04-4**

Beginning RT Labcode **RT1267**

Ending RT Labcode **RT1267**

3:12:50PM

12/27/2004